

8/20/18

Apex Environmental
P.O. Box 1445
Wilsonville, OR 97070

Re: Lewis Elementery

Dear Apex Environmental

Enclosed are the results of analysis for samples received by the laboraory on 8/20/2018
The results related only to the samples included in this report.
The project was assigned a report number of 89483

If you have any questions concerning this report, please feel free to contact us.

Sincerely,



CY Chan
QA Officer

Laboratory Report

Apex Environmental
 P.O. Box 1445
 Wilsonville, OR 97070

Project Name: Lewis Elementary
 Project Location:
 Project Number:
 Date Sampled:
 Date received: 8/20/18

Report Number: 89483
 Report Date: 8/20/18

NIOSH 7105

Analyte: Total Lead (Pb) in wipe quantitation

Field ID	Lab ID	Result (µg/sq ft)	Area Wiped (sq ft)	Detection Limit (µg/sq ft)
LLW-01	N7859	< 29	1	29
LLW-02	N7860	< 29	1	29
LLW-03	N7861	< 29	1	29
LLW-04	N7862	< 31	1	31
LLW-05	N7863	< 30	1	30
LLW-06	N7864	< 30	1	30
LLW-07	N7865	< 29	1	29

Quality Control Report: Lead (Pb)

	Measured Conc. 1 (mg/L)	Theoretical Concentration (mg/L)	Recovery (%)	Lower Limit	Upper Limit	
ICV	5	5	95%	80%	120%	PASS
Prep Blank	0	0	--	0.5 mg/L	--	PASS
LCS	3	3	98%	70%	130%	PASS
QC Blank	0	0	--	0.5 mg/L	--	PASS
QC Check	5	5	103%	80%	120%	PASS
QC Blank	0	0	--	0.5 mg/L	--	PASS
QC Check	5	5	101%	80%	120%	PASS
QC Blank	0	0	--	0.5 mg/L	--	PASS
QC Check	5	5	106%	80%	120%	PASS
QC Blank	0	0	--	0.5 mg/L	--	PASS
QC Check	5	5	105%	80%	120%	PASS
QC Blank	0	0	--	0.5 mg/L	--	PASS
QC Check	5	5	104%	80%	120%	PASS
Calibration 1.		(Lower Limit: 0.995)				PASS



Environmental Sciences, Inc.
2415 SE 11th Ave. Portland Oregon 97214

CHAIN OF CUSTODY

Report Number 89483

AD/QE-0017

Phone(503) 231-9320 FAX(503) 231-9344

Company NEX Environmental		Phone 503 789 4185		FAX									
Project #		Purchase Order #		Report Attention									
Project Name Lewis elementary		Report Attention		Collected By									
Site		Turnaround Time: <input type="checkbox"/> Regular <input type="checkbox"/> 3-5 Business Days											
Samples: Temperature _____ On Ice? Yes / No													
LAB ID	Field ID	Sampling Date	Sampling Time	Matrix	Container	Volume	NW-TPH-Dx	NW-TPH-GX	NW-TPH-HCID	EPA 8021B (BTEX)	EPA 8270 SIM (PAH)	EPA 8260B	Comments
N7859	LLW -01	8/20	10:30		240ml	150ff							lead pipe
N7860	-02												
N7861	-03												
N7862	-04												
N7863	-05												
N7864	-06												
N7865	-07												
Relinquished by <i>[Signature]</i>		Affiliation AEX	Date 8/20	Time 2:11	Received by <i>[Signature]</i>	Affiliation WV East	Date 8/20/18	Time 14:14					
Relinquished by		Affiliation	Date	Time	Received by	Affiliation	Date	Time					